

ABX Minotrol CRP



LOT

MC 431

CONTROL



(Exp.) **2024-07-05**

(YYYY - MM - DD)

Rev 2

		ABX Lyse																
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3		
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE		
				60 8P 60 18P					60 8P 60 18P					60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l			2.1		± 0.4			7.9			± 1.0			21.1			± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l			2.25		± 0.15			4.54			± 0.18			5.59			± 0.25
HB HGB	g/dl			6.2		± 0.4			13.4			± 0.5			17.7			± 0.7
	g/l			62		± 4			134			± 5			177			± 7
	mmol/l			3.85		± 0.25			8.32			± 0.31			10.99			± 0.44
HT HCT	%			16.7		± 2.0			36.8			± 2.5			49.8			± 3.0
	l/l			0.167		± 0.020			0.368			± 0.025			0.498			± 0.030
VGM MCV	µm ³ ; fl			74		± 4			81			± 4			89			± 4
TGMH MCH	pg			27.6		± 2.4			29.5			± 2.4			31.7			± 2.8
	fmol			1.71		± 0.15			1.83			± 0.15			1.97			± 0.18
CCMH MCHC	g/dl			37.2		± 3.0			36.4			± 3.0			35.6			± 3.0
	g/l			372		± 30			364			± 30			356			± 30
	mmol/l			23.12		± 1.86			22.63			± 1.86			22.09			± 1.86
IDR RDW	%			14.0		± 3.0			14.0			± 3.0			13.0			± 3.0
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l			80		± 20			268			± 40			532			± 60
VPM MPV	µm ³ ; fl			7.5		± 2.0			6.8			± 2.0			7.1			± 2.0
LYMPHO	#			1.3		± 0.2			2.7			± 0.5			3.3			± 1.1
	%			64.0		± 8.0			34.0			± 6.0			15.5			± 5.0
MONO	#			0.2		± 0.2			1.2			± 1.1			1.7			± 1.3
	%			10.0		± 7.0			15.3			± 13.8			8.0			± 6.0
GRANULO	#			0.5		± 0.2			4.0			± 1.0			16.1			± 2.1
	%			26.0		± 7.0			50.7			± 10.0			76.5			± 10.0
CRP	mg/dl			N/A		± 0.30			N/A			± 0.70			N/A			± 1.00
	mg/l			N/A		± 3.0			N/A			± 7.0			N/A			± 10.0

Ref: TEMP-1070 Rev.42 FRONT/RECTO 1300092170